

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-01292024-005

Gage ID PEA-28044 Gage S/N 28044 Description 10 lb Test Weight Operating Procedure: OP-118 Unit of Meas. pounds Manufacturer Parker Cal. Date 1/29/2024 Next Due 1/29/2027 Cal. Freq. 3.00 Years Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 12519 Job 12299

Using calibrated scale serial no. 8335029565

Test weight as found 10.060 lbs.


Test weight as left 10.060 lbs.

Calibration due date for scale is on 7/26/2024

Scale is traceable to NIST by the following numbers: Doc# 138436

Calibrated By Chad Adams

Signature

Date: 1/29/2024**APPROVED**

By Kayla Myers at 11:31 am, Feb 03, 2024