



Tubular Inspection Products Inc.

Certificate of Calibration

This Calibration Certificate confirms that the instruments described herein have been calibrated in compliance with the requirements set forth by ISO/IEC 17025:2017, ensuring that all measurement results are traceable to national or international standards, including those established by the National Institute of Standards and Technology (NIST). This certificate reflects adherence to the manufacturer's specifications, demonstrating compliance with established guidelines for performance and reliability.

Standards: ISO17025:2017/ISO9001/API Q1/7-2

Customer: **Peak NDT Solutions**

Calibration Date: **July 13, 2026**

Calibration Due Date: **January 13, 2027**

Serial Number: **22953**

Instrument: **ZX-3**

Manufacturer: **Dakota NDT**

TIP Procedure: **TIP-UTW01.1**

Calibration Technician: **Ryan Britton**

Order Number: **80910**

Results

Reference Size	Actual Size	Deviation	Tolerance	Result
<u>0.100" Step</u>	<u>0.099"</u>	<u>0.001"</u>	<u>+/-0.002"</u>	<u>Pass</u>
<u>0.199" Step</u>	<u>0.200"</u>	<u>0.001"</u>	<u>+/-0.002"</u>	<u>Pass</u>
<u>0.299" Step</u>	<u>0.300"</u>	<u>0.001"</u>	<u>+/-0.002"</u>	<u>Pass</u>
<u>0.399" Step</u>	<u>0.400"</u>	<u>0.001"</u>	<u>+/-0.002"</u>	<u>Pass</u>
<u>0.499" Step</u>	<u>0.500"</u>	<u>0.001"</u>	<u>+/-0.002"</u>	<u>Pass</u>

<u>Instrument Received</u>	<u>Instrument Returned</u>	<u>Instrument Status</u>
In Tolerance	In Tolerance	Passed

Temperature: 68.5°F, RH 40.5%

APPROVED

By Kayla Myers at 8:52 am, Jul 24, 2026

Issue Date for Certificate: **July 13, 2026**

TIP Master Optical Comparator, Keyence IM-8030

SN: 4C0Q000412 Cal Date: 06-26-2027

TIP -MS-SB SN: K08133 Cal Date: 01-29-27

TIP Master Hygrometer SN: 31874/1H2048MCR, 130070751, Cal Date: 01-25-27

Approved By: 