

TSP Calibration, Inc.

1110 S. FM 1788 Suite 38 Midland, TX 79706
Phone: (432) 296-4992 Email: Chad@tspndt.com

Certificate of Calibration

Certificate No TSPW-07222025-005

Gage ID PEA-222500427 Gage S/N 222500427 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 7/22/2025 Next Due 1/22/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 14594 Job 14235

Calibrated black light serial no. 1911981 was used during this calibration.

Calibration due date for black light is on 8/8/2025

Black light is traceable to NIST by the following numbers: Doc# 154779

Standard ID	12"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2912.7	Nominal	3066	Maximum	3219.3
Ref Type	Light Meter	As Found	3108	Accuracy	42	Fail Before	No
		As Left	3108	Accuracy	42	Fail After	No
Gage ID of Standard	REF-1705427						
Std Due Date	8/12/2025	Gage S/N	1705427	NIST No.	154789		
Standard ID	15"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2006.4	Nominal	2112	Maximum	2217.6
Ref Type	Light Meter	As Found	2146	Accuracy	34	Fail Before	No
		As Left	2146	Accuracy	34	Fail After	No
Gage ID of Standard	REF-1705427						
Std Due Date	8/12/2025	Gage S/N	1705427	NIST No.	154789		

Calibrated By Chad Adams

Signature



Date: 7/22/2025

APPROVED

By Kayla Myers at 8:15 am, Jul 28, 2025