

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-01302026-013

<p>Gage ID PEA-222500427 Gage S/N 222500427 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 1/30/2026 Next Due 7/30/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings


Job 15136

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
 Reference Standard: UV Light - Serial Number: 1911981
 NIST Number: Doc#171852 - Calibration Due Date: 8/21/2026

Standard ID	12"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2896.55	Nominal	3049	Maximum	3201.45
Ref Type	Light Meter	As Found	3116	Accuracy	67	Fail Before	No
		As Left	3116	Accuracy	67	Fail After	No
Gage ID of Standard	REF-1705427	Gage S/N	1705427	NIST No.	171854		
Std Due Date	8/21/2026						

Standard ID	15"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2179.3	Nominal	2294	Maximum	2408.7
Ref Type	Light Meter	As Found	2329	Accuracy	35	Fail Before	No
		As Left	2329	Accuracy	35	Fail After	No
Gage ID of Standard	REF-1705427	Gage S/N	1705427	NIST No.	171854		
Std Due Date	8/21/2026						

Calibrated By Chad Adams

Signature 

Date: 1/30/2026

APPROVED
By Kayla Myers at 3:34 pm, Feb 06, 2026