

## TSP Calibration, Inc.

1110 S. FM 1788 #42  
 Phone: (432) 296-4992 Email: Chad@tspndt.com  
 Midland, TX 79706

### Certificate of Calibration

Certificate No TSPW-01302026-014

<p><b>Gage ID</b> PEA-220100735  <b>Gage S/N</b> 220100735  <b>Description</b> White Light Meter  <b>Operating Procedure:</b> OP-130  <b>Unit of Meas.</b> Foot Candles  <b>Manufacturer</b> General  <b>Cal. Date</b> 1/30/2026  <b>Next Due</b> 7/30/2026  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 Deg F +/- 2                  Humidity 35-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
---	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

Job 15136

	Master	As Found
High:	102.3	101.9
Low:	16.8	15.4

NOTE: IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND" AND "AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS.

Tolerance: +/-5% or +/-2 foot candles, whichever is greater.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):  
 Reference Standard: Dual Light Meter - Serial Number: 1705427 Sensor: 1705430  
 NIST Number: Doc#171854 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams Signature  Date: 1/30/2026

**APPROVED**  
 By Kayla Myers at 3:59 pm, Feb 06, 2026