TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-10192023-014

Gage ID PEA-28038
Gage S/N 28038

Description 10 lb. Test Weight

Operating OP-118

Procedure: Unit of Meas. Ibs

Manufacturer Parker

Cal. Date 10/19/2023 Next Due 10/19/2026

Cal. Freq. 3.00 Years

Location Lab

Environmental Conditions

Temperature 68 +/- 2 deg F

Humidity 20-55%

Approved Yes

Customer Info. Peak NDT

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 12225 JOB 12846

AS FOUND: PASSED AS LEFT: PASSED

Using calibrated scale serial no. 7129341906

Test weight as found 10.08 lbs. Test weight as left 10.08 lbs.

Calibration due date for scale is on 4/13/2024.

Scale is traceable to NIST by the following numbers:LA22-255, LA22-256, LA22-257

Calibrated By Taichi Daimo Signature Date: 10/19/2023

APPROVED

By Kayla Myers at 4:04 pm, Oct 19, 2023