

Tech Service Products

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-10132026-011

Gage ID PEA-K03175 Gage S/N K03175 Description 5 Step Block Operating Procedure: OP-126 Unit of Meas. Inch Manufacturer N/A Cal. Date 10/13/2026 Next Due 10/13/2027 Cal. Freq. 1.00 Years Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 20-55% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No. 1066378 Job No. 15301

Standard ID	0.100"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.098	Nominal	0.100	Maximum	0.102
Ref Type	12" Digital Caliper	As Found	0.100	Accuracy	0.000	Fail Before	No
		As Left	0.100	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-1095377						
Std Due Date	8/27/2027	Gage S/N	1095377	NIST No.	187830		
Standard ID	0.200"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.198	Nominal	0.200	Maximum	0.202
Ref Type	12" Digital Caliper	As Found	0.200	Accuracy	0.000	Fail Before	No
		As Left	0.200	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-1095377						
Std Due Date	8/27/2027	Gage S/N	1095377	NIST No.	187830		
Standard ID	0.300"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.298	Nominal	0.300	Maximum	0.302

Ref Type	12" Digital Caliper	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date 8/27/2027

Gage S/N 1095377

NIST No. 187830

Standard ID	0.400"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	12" Digital Caliper	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date 8/27/2027

Gage S/N 1095377

NIST No. 187830

Standard ID	0.500"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	12" Digital Caliper	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-1095377

Std Due Date 8/27/2027

Gage S/N 1095377

NIST No. 187830

Calibrated By Chad Adams

Signature



Date: 10/13/2026

APPROVED

By Kayla Myers at 9:25 am, Oct 14, 2026