

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-04152026-000

<p><b>Gage ID</b> PEA-14669  <b>Gage S/N</b> 14669  <b>Description</b> A/C Yoke  <b>Operating Procedure:</b> OP-115  <b>Unit of Meas.</b> lbs  <b>Manufacturer</b> Magnaflux  <b>Cal. Date</b> 4/15/2026  <b>Next Due</b> 10/15/2026  <b>Cal. Freq.</b> 6.00                      Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b></p> <p>Temperature    68 +/- 2 deg F                  Humidity            20-55%</p> <p style="text-align: center;"><b>Approved</b>    Yes  <b>Customer Info.</b>    Peak NDT</p>
---	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

AS FOUND: PASSED      AS LEFT: PASSED

A/C Yoke picked up and held required 10 lb. test weight for calibration.  
 Calibrated test weight serial no. 2768 was used for this calibration.  
 Test weight calibration is due on 11/19/2028.  
 Test weight is traceable to NIST by the following numbers: 25-334, 25-376, 25-335

Calibrated By   Taichi Daimo        Signature           Date:   4/15/2026  

**APPROVED**  
 By Kayla Myers at 2:02 pm, Apr 15, 2026