

Tech Service Products

1110 S. FM 1788 #42

Phone: (432) 296-4992 Email: Chad@tspndt.com

Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-10132026-009

Gage ID PEA-0209221592 Gage S/N 0209221592 Description 12" Digital Depth Gauge Operating Procedure: OP-123 Unit of Meas. Inch Manufacturer Insize Cal. Date 10/13/2026 Next Due 4/13/2027 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 20-55% Approved Yes Customer Info. Peak NDT Solutions
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No. 1066378 Job No. 15301

Standard ID	1.000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.999	Nominal	1.000	Maximum	1.001
Ref Type	81PC Gage Blocks	As Found	0.999	Accuracy	-0.001	Fail Before	No
		As Left	0.999	Accuracy	-0.001	Fail After	No
Gage ID of Standard	REF-070111	Gage S/N	070111	NIST No.	81004		
Std Due Date	1/26/2027						
Standard ID	3.000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	2.999	Nominal	3.000	Maximum	3.001
Ref Type	81PC Gage Blocks	As Found	3.000	Accuracy	0.000	Fail Before	No
		As Left	3.000	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-070111	Gage S/N	070111	NIST No.	81004		
Std Due Date	1/26/2027						
Standard ID	5.000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	4.999	Nominal	5.000	Maximum	5.001

Ref Type	81PC Gage Blocks	As Found	5.000	Accuracy	0.000	Fail Before	No
		As Left	5.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Standard ID	7.000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	81PC Gage Blocks	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111

Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Standard ID	9.000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	8.999	Nominal	9.000	Maximum	9.001
Ref Type	81PC Gage Blocks	As Found	9.000	Accuracy	0.000	Fail Before	No
		As Left	9.000	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-070111


Std Due Date 1/26/2027

Gage S/N 070111

NIST No. 81004

Calibrated By Chad Adams

Signature



Date: 10/13/2026

APPROVED

By Kayla Myers at 9:21 am, Oct 14, 2026