Tech Service Products

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland,TX 79706

Certificate of Calibration

Certificate No TSPW-10132026-007

Gage ID	PEA-12846
Gage S/N	12846
Description	20-0-20 Gauss Meter
Operating Procedure:	OP-117
Unit of Meas.	Gauss
Manufacturer	R.B. Annis
Cal. Date	10/13/2026
Next Due	4/13/2027
Cal. Freq.	6.00 Months
Location	Lab

Environmental Conditions

Temperature	68 Deg F +/- 2
Humidity	20-55%

Approved Yes
Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No. 1066378 Job No. 15301

Standard ID	+10	**Uncertainty**			**Units**		**Type**	V	
Limited Use?	No	**Minimum**	9	**Nominal**	10	**Maximum**	11		
Ref Type	20-0-20 Gauss Meter	**As Found**	10	**Accuracy**	0	**Fail Before**	No		
		As Left	10	**Accuracy**	0	**Fail After**	No		
Gage ID of Standard	REF-TSPSTD-02								
Std Due Date	8/27/2027	**Gage S/N**	TSPSTD-02	**NIST No.**	187829				

Standard ID	+20	**Uncertainty**			**Units**		**Type**	V	
Limited Use?	No	**Minimum**	19	**Nominal**	20	**Maximum**	21		
Ref Type	20-0-20 Gauss Meter	**As Found**	20	**Accuracy**	0	**Fail Before**	No		
		As Left	20	**Accuracy**	0	**Fail After**	No		
Gage ID of Standard	REF-TSPSTD-02								
Std Due Date	8/27/2027	**Gage S/N**	TSPSTD-02	**NIST No.**	187829				

Standard ID	-10	**Uncertainty**			**Units**		**Type**	V	
Limited Use?	No	**Minimum**	9	**Nominal**	10	**Maximum**	11		

Gage ID: PEA-12846

Ref Type	20-0-20 Gauss Meter	**As Found**	10	**Accuracy**	0	**Fail Before**	No
		As Left	10	**Accuracy**	0	**Fail After**	No

Gage ID of Standard	REF-TSPSTD-02				
Std Due Date	8/27/2027	**Gage S/N**	TSPSTD-02	**NIST No.**	187829

Standard ID	-20	**Uncertainty**		**Units**		**Type**	V
Limited Use?	No	**Minimum**	19	**Nominal**	20	**Maximum**	21
Ref Type	20-0-20 Gauss Meter	**As Found**	20	**Accuracy**	0	**Fail Before**	No
		As Left	20	**Accuracy**	0	**Fail After**	No

Gage ID of Standard	REF-TSPSTD-02				
Std Due Date	8/27/2027	**Gage S/N**	TSPSTD-02	**NIST No.**	187829

Calibrated By Chad Adams **Signature** _____ **Date:** 10/13/2026

APPROVED

By Kayla Myers at 9:17 am, Oct 14, 2026