Gage ID: PEA-12846 Page 1 of 2

# TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

## **Certificate of Calibration**

#### Certificate No TSP-10062025-003

Gage ID PEA-12846

Gage S/N 12846 **Description** 20-0-20 Gauss Meter

Operating OP-117

Procedure:

Unit of Meas. Gauss

Manufacturer R. B. Annis Co., Inc.

Cal. Date 10/6/2025 Next Due 4/6/2026

Cal. Freq. 6.00

Location Lab

Months

**Environmental Conditions** 

**Temperature** 68 +/- 2 deg F

Humidity 20-55%

> **Approved** Yes Customer Info. Peak NDT

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### **Findings**

Standard ID	+10		Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum		
Ref Type	Gauss	Meter	As Found		10	Accuracy		0	Fail Before	No	
<b>,</b> ,			As Left		10	Accuracy		0	Fail After	No	
Gage ID of Star	ndard	Ref-09-1229									
Std Due Date		10/10/2025	Gage S/N	09-1229	NIST No.		114537				
Standard ID	+20		Uncertainty			Units			Туре	V	_
Limited Use?	No		Minimum		19	Nominal		20	Maximum	21	
Ref Type	Gauss	Meter	As Found		20	Accuracy		0	Fail Before	No	
			As Left		20	Accuracy		0	Fail After	No	
Gage ID of Standard Ref-09-1229											
Std Due Date		10/10/2025	Gage S/N	09-1229	NIST No.		114537				
Standard ID	-10		Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	

Ref Type Gauss Meter As Found 10 Accuracy Fail Before No 0 As Left 10 Accuracy Fail After No Gage ID of Standard Ref-09-1229 Std Due Date 10/10/2025 Gage S/N 09-1229 NIST No. 114537 Standard ID Type V Uncertainty Units Limited Use? No Minimum 19 Nominal 20 Maximum 21 Ref Type As Found 0 Gauss Meter 20 Fail Before No Accuracy As Left Fail After No 20 Accuracy Gage ID of Standard Ref-09-1229 Std Due Date 10/10/2025 Gage S/N 09-1229 NIST No. 114537

Page

Date:

2 of 2

10/6/2025

Calibrated By Taichi Daimo Signature

Gage ID: PEA-12846

APPROVED

By Kayla Myers at 2:07 pm, Oct 07, 2025