Tech Service Products

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland,TX 79706

Certificate of Calibration

Certificate No TSPW-11022026-002

Gage ID	PEA-7339
Gage S/N	7339
Description	Pit Gauge
Operating Procedure:	OP-122
Unit of Meas.	Inch
Manufacturer	Gagemaker
Cal. Date	11/2/2026
Next Due	11/2/2027
Cal. Freq.	1.00 Years
Location	Lab

Environmental Conditions

Temperature	68 Deg F +/- 2
Humidity	20-55%

Approved	Yes
Customer Info.	Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No. 1067815 Job No. 15317

Standard ID	0.100"	**Uncertainty**			**Units**			**Type**	V
Limited Use?	No	**Minimum**		0.099	**Nominal**		0.100	**Maximum**	0.101
Ref Type	Bench Calibrator	**As Found**		0.100	**Accuracy**		0.000	**Fail Before**	No
		As Left		0.100	**Accuracy**		0.000	**Fail After**	No
Gage ID of Standard	REF-I132869								
Std Due Date	8/21/2027	**Gage S/N**	I132869		**NIST No.**	89114			

Standard ID	0.200"	**Uncertainty**			**Units**			**Type**	V
Limited Use?	No	**Minimum**		0.199	**Nominal**		0.200	**Maximum**	0.201
Ref Type	Bench Calibrator	**As Found**		0.200	**Accuracy**		0.000	**Fail Before**	No
		As Left		0.200	**Accuracy**		0.000	**Fail After**	No
Gage ID of Standard	REF-I132869								
Std Due Date	8/21/2027	**Gage S/N**	I132869		**NIST No.**	89114			

Standard ID	0.300"	**Uncertainty**			**Units**			**Type**	V
Limited Use?	No	**Minimum**		0.299	**Nominal**		0.300	**Maximum**	0.301

Ref Type	Bench Calibrator	**As Found**	0.300	**Accuracy**	0.000	**Fail Before**	No
		As Left	0.300	**Accuracy**	0.000	**Fail After**	No
Gage ID of Standard	REF-I132869						
Std Due Date	8/21/2027	**Gage S/N**	I132869	**NIST No.**	89114		

Standard ID	0.400"	**Uncertainty**		**Units**		**Type**	V
Limited Use?	No	**Minimum**	0.399	**Nominal**	0.400	**Maximum**	0.401
Ref Type	Bench Calibrator	**As Found**	0.400	**Accuracy**	0.000	**Fail Before**	No
		As Left	0.400	**Accuracy**	0.000	**Fail After**	No
Gage ID of Standard	REF-I132869						
Std Due Date	8/21/2027	**Gage S/N**	I132869	**NIST No.**	89114		

Standard ID	0.500"	**Uncertainty**		**Units**		**Type**	V
Limited Use?	No	**Minimum**	0.499	**Nominal**	0.500	**Maximum**	0.501
Ref Type	Bench Calibrator	**As Found**	0.500	**Accuracy**	0.000	**Fail Before**	No
		As Left	0.500	**Accuracy**	0.000	**Fail After**	No
Gage ID of Standard	REF-I132869						
Std Due Date	8/21/2027	**Gage S/N**	I132869	**NIST No.**	89114		

Calibrated By Chad Adams **Signature**  **Date:** 11/2/2026

APPROVED
By Kayla Myers at 10:18 pm, Nov 03, 2026