

Tech Service Products

1110 S. FM 1788 #42

Phone: (432) 296-4992 Email: Chad@tspndt.com

Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-10132026-010

Gage ID PEA-23040 Gage S/N 23040 Description UT Thickness Gauge Operating Procedure: OP-121 Unit of Meas. Inch Manufacturer Dakota Cal. Date 10/13/2026 Next Due 4/13/2027 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 20-55% Approved Yes Customer Info. Peak NDT Solutions
---	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No. 1066378 Job No. 15301

Standard ID	0.100"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.098	Nominal	0.100	Maximum	0.102
Ref Type	5 Step Block	As Found	0.100	Accuracy	0.000	Fail Before	No
		As Left	0.100	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-337214						
Std Due Date	8/27/2027	Gage S/N	337214	NIST No.	187834		
Standard ID	0.200"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.198	Nominal	0.200	Maximum	0.202
Ref Type	5 Step Block	As Found	0.200	Accuracy	0.000	Fail Before	No
		As Left	0.200	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-337214						
Std Due Date	8/27/2027	Gage S/N	337214	NIST No.	187834		
Standard ID	0.300"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.298	Nominal	0.300	Maximum	0.302

Gage ID: PEA-23040

Page 2 of 2

Ref Type	5 Step Block	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214
Std Due Date 8/27/2027

Gage S/N 337214 NIST No. 187834

Standard ID	0.400"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	5 Step Block	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214
Std Due Date 8/27/2027

Gage S/N 337214 NIST No. 187834


Standard ID	0.500"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	5 Step Block	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-337214
Std Due Date 8/27/2027

Gage S/N 337214 NIST No. 187834

Calibrated By Chad Adams

Signature



Date: 10/13/2026

APPROVED

By Kayla Myers at 9:24 am, Oct 14, 2026