

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-04072026-006

<p><b>Gage ID</b> PEA-2240000210  <b>Gage S/N</b> 2240000210  <b>Description</b> UV Light Meter  <b>Operating Procedure:</b> OP-119  <b>Unit of Meas.</b> Microwatts/cm2  <b>Manufacturer</b> General  <b>Cal. Date</b> 4/7/2026  <b>Next Due</b> 10/7/2026  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT</p>
---	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

Calibrated black light serial no. 2234240 was also used during this calibration.  
 Calibration due date for black light is on 06/17/2026

Black light is traceable to NIST by the following numbers: 263546, 263547

<b>Standard ID</b> 12"	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b> 6485	<b>Nominal</b> 6826	<b>Maximum</b> 7167
<b>Ref Type</b> Long Wave	<b>As Found</b> 6988	<b>Accuracy</b> 162	<b>Fail Before</b> No
	<b>As Left</b> 6988	<b>Accuracy</b> 162	<b>Fail After</b> No
<b>Gage ID of Standard</b> REF-1944278	<b>Gage S/N</b> 1944278/1944	<b>NIST No.</b> 263546, 263547	
<b>Std Due Date</b> 6/17/2026	280		

<b>Standard ID</b> 15"	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b> 4476	<b>Nominal</b> 4711	<b>Maximum</b> 4946
<b>Ref Type</b> Long Wave	<b>As Found</b> 4872	<b>Accuracy</b> 161	<b>Fail Before</b> No
	<b>As Left</b> 4872	<b>Accuracy</b> 161	<b>Fail After</b> No
<b>Gage ID of Standard</b> REF-1944278	<b>Gage S/N</b> 1944278/1944	<b>NIST No.</b> 263546, 263547	
<b>Std Due Date</b> 6/17/2026	280		

Calibrated By Taichi Daimo

Signature



Date: 4/7/2026

**APPROVED**  
*By Kayla Myers at 8:18 am, Apr 09, 2026*