TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-04082025-024

Gage ID PEA-220702539 Gage S/N 220702539

Description Visible Light Meter

Operating OP-130 Procedure:

Unit of Meas. Foot Candles

Manufacturer General Cal. Date 4/8/2025

Next Due 10/8/2025

Cal. Freq. 6.00

Months

Location Lab

Environmental Conditions

Temperature

68 +/- 2 deg F

Humidity

20-55%

Approved Yes Customer Info. Peak NDT

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Using calibrated light meter serial no. 1944278 with sensor 1944280 the following readings were taken.

Readings As Found (foot candles)

Master Actual

High 96.5 94.7 33.2 31.8 Low

Readings As Left (foot candles)

Master Actual High 96.5 94.7 Low 33.2 31.8

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater

Cal due date for light meter is 06/17/2025

NIST Traceable through: 258648, 258649, 258650

APPROVED

By Kayla Myers at 3:14 pm, Apr 09, 2025

Signature (Calibrated By Taichi Daimo 4/8/2025