

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-04072026-005

Gage ID PEA-220702539 Gage S/N 220702539 Description Visible Light Meter Operating Procedure: OP-130 Unit of Meas. Foot Candles Manufacturer General Cal. Date 4/7/2026 Next Due 10/7/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Using calibrated light meter serial no. 1944278 with sensor 1944280 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	96.5	94.3
Low	35.3	33.6

Readings As Left (foot candles)

	Master	Actual
High	96.5	94.3
Low	35.3	33.6

APPROVED
 By Kayla Myers at 8:17 am, Apr 09, 2026

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater
 Cal due date for light meter is 06/17/2026
 NIST Traceable through : 263546, 263547

Calibrated By Taichi Daimo Signature  Date: 4/7/2026