Tech Service Products

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland,TX 79706

Certificate of Calibration

Certificate No TSPW-10132026-005

Gage ID	PEA-220702539
Gage S/N	220702539
Description	White Light Meter
Operating Procedure:	OP-130
Unit of Meas.	Foot Candles
Manufacturer	General
Cal. Date	10/13/2026
Next Due	4/13/2027
Cal. Freq.	6.00 Months
Location	Lab

Environmental Conditions

Temperature	68 Deg F +/- 2
Humidity	20-55%

Approved	Yes
Customer Info.	Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Order No. 1066378 Job No. 15301

As Found: Passed As Left: Passed

	Master	As Found
High:	126.3	122.1
Low:	13.8	12.0

NOTE: If only an "As Found" reading is stated, "As Found" and "As Left" are the same with no adjustments.

Tolerance: +/-5% or +/-2 foot candles, whichever is greater.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):
Reference Standard: Dual Light Meter - Serial Number: 1705427 Sensor: 1705430
NIST Number: 187827 - Calibration Due Date: 8/27/2027

Calibrated By Chad Adams **Signature** _____ **Date:** 10/13/2026

APPROVED
By Kayla Myers at 9:15 am, Oct 14, 2026