

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-05052026-016

<p>Gage ID PEA-1896663 Gage S/N 1896663/1896673 Description Dual Light Meter Operating Procedure: OP-119/OP-130 Unit of Meas. Microwatts/cm2 & Foot Candles Manufacturer Spectroline Cal. Date 5/5/2026 Next Due 11/5/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Light Meter s/n: 1896663
 Sensor s/n: 1896673

Using calibrated light meter serial no. 1944278 with sensor 1944280 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	96.4	94.1
Low	35.3	33.6

Readings As Left (foot candles)

	Master	Actual
High	96.4	94.1
Low	35.3	33.6

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater
 Cal due date for light meter is 06/17/2026
 NIST Traceable through : 263546, 263547

Calibrated black light serial no. 2234240 was also used during this calibration.
 Calibration due date for black light is on 06/17/2026
 Black light is traceable to NIST by the following numbers: 263546, 263547

Standard ID	12"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6537	Nominal	6881	Maximum	7225
Ref Type	Long Wave	As Found	6733	Accuracy	-148	Fail Before	No
		As Left	6733	Accuracy	-148	Fail After	No

Gage ID of Standard REF-1944278

Std Due Date 6/17/2026

Gage S/N 1944278/1944 NIST No. 263546, 263547
280

Standard ID	15"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	4391	Nominal	4622	Maximum	4853
Ref Type	Long Wave	As Found	4470	Accuracy	-152	Fail Before	No
		As Left	4470	Accuracy	-152	Fail After	No

Gage ID of Standard REF-1944278

Std Due Date 6/17/2026

Gage S/N 1944278/1944 NIST No. 263546, 263547
280

Calibrated By Taichi Daimo Signature  Date: 5/5/2026

APPROVED
By Kayla Myers at 8:05 am, May 08, 2026