



CALIBRATION CERTIFICATION

INSTRUMENT: MIZ-21C EDDY CURRENT TESTER

SERIAL NUMBER: 794722

CONDITION CODE: B

The calibration of this instrument has been performed using a documented, controlled Zetec procedure which meets or exceeds the manufacturer's requirements.

CUSTOMER: Peak NDT

515a Verot School Rd Lafayette, LA 70508 US

Zetec Calibration Laboratories have been audited and found in compliance with ISO/IEC 17025:2017. Accredited calibrations performed within the Lab's Scope of Accreditation are indicated by the presence of the Accrediting Body's name and Zetec's Certification Number on this Certificate of Calibration. Zetec hereby certifies that the equipment calibrated per this certificate meets or exceeds all manufacture specifications and has been calibrated using a documented, controlled Zetec procedure accomplished in compliance with the current Zetec Quality Program. This calibration has been performed using measurement standards whose accuracies are traceable to the International System of Units. The certificate results relate only to this calibrated instrument. There is no implied warranty that the instrument will maintain its specified tolerances during the calibration interval due to possible drift, environment, or other factors beyond Zetec control. Unless otherwise specified, Zetec utilizes the simple decision rule based on tolerance boundaries. All purchase order requirements have been met. This certificate may not be reproduced or distributed except in whole without the written authorization from Zetec.

N.I.S.T TRACEABILITY:

Standard Instrument		Serial Number	Calibration Date	Expiration
4853	FUNCTION GENERATOR	MY40011156	04-21-2023	04-21-2024
7682	DIGITAL MULTIMETER	MY45003222	02-07-2023	02-07-2024
AMS152	TEMP/HUMIDITY LOGGER	10467293	04-28-2023	04-28-2024

CALIBRATION DATE:

29 December 2023

CALIBRATED BY: JAMES LARSEN/SR. CALIBRATION TECH

EXPIRATION DATE:

29 December 2024

REVIEWED BY: TRAVIS SATTERLEE/SYSTEMS DELIVERY LEAD

Procedure#: SLP-3114-PR-Z REV. C

WO#: 23393

COMMENTS:

TEMPERATURE: 21.1 DEG C HUMIDITY: 23.6% RH
DIGITAL READOUT IN 0.9 TO 110% IACS (0.5 TO 70 MS/M), ACCURACY WITHIN +/-0.5% IACS FROM 0.9% TO 65% IACS AND WITHIN +/-1.0% OF VALUES OVER 65%
MIZ-21C SINGLE FREQUENCY SEALED
SOFTWARE VERSION: 1.4.1.24
FFU: 1.2.0.19
FPGA: 1.50

APPROVED**By Kayla Myers at 3:16 pm, Mar 27, 2024**

Calibration Certification Number: 8119



SERVICE REPORT

Zetec, Inc.,
8226 Bracken Place SE, Suite 100
Snoqualmie, WA 98065 U.S.A.
(425) 974-2700 Fax (425) 974-2701
Orders: customerservice@zetec.com
www.zetec.com

Printed Date (mm/dd/yyyy) 12/28/2023	Work Order # 23393	Page 1 / 1
Customer # 8295	Customer Name Peak NDT	
Customer PO's No.	Zetec Customer Order Reference M6913	Order Type Service Order

Customer Object ID: 111A901-00-794722
Description: ZES-HHT-MIZ-21C-SF

Request: MIZ-21C SF s/n 794722

Received Condition: WITHIN TOLERANCES
Service Type: REGULAR MAINTENANCE

Cause: Calibration/Alignment
Priority: Normal Turnaround Expectations

Cause Description: STANDARD CALIBRATION REQUESTED

Services Performed:
CALIBRATION AND CERTIFICATION
CALIBRATED AND FUNCTION TESTED

Technician: JAMESL
Date Completed: 12/28/2023

Zetec Facility: ZETEC Snoqualmie S354