

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-05052026-018

<p>Gage ID PEA-BR11GM0038 Gage S/N BR11GM0038 Description Internal Taper Gage Operating Procedure: OP-122 Unit of Meas. inch Manufacturer Gagemaker Cal. Date 5/5/2026 Next Due 11/5/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID .10000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.09900	Nominal 0.10000	Maximum 0.10100
Ref Type Mic Trac	As Found 0.10005	Accuracy 0.00005	Fail Before No
	As Left 0.10005	Accuracy 0.00005	Fail After No
Gage ID of Standard REF-G141308	Gage S/N G141308	NIST No. 747V66-240502-00	
Std Due Date 10/1/2026			

Standard ID .20000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.19900	Nominal 0.20000	Maximum 0.20100
Ref Type Mic Trac	As Found 0.20007	Accuracy 0.00007	Fail Before No
	As Left 0.20007	Accuracy 0.00007	Fail After No
Gage ID of Standard REF-G141308	Gage S/N G141308	NIST No. 747V66-240502-00	
Std Due Date 10/1/2026			

Standard ID .30000 "	Uncertainty	Units	Type V
Limited Use? No	Minimum 0.29900	Nominal 0.30000	Maximum 0.30100

Ref Type	Mic Trac	As Found	0.30005	Accuracy	0.00005	Fail Before	No
		As Left	0.30005	Accuracy	0.00005	Fail After	No

Gage ID of Standard	REF-G141308	Gage S/N	G141308	NIST No.	747V66-240502-00
Std Due Date	10/1/2026				

Standard ID	.40000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.39900	Nominal	0.40000	Maximum	0.40100
Ref Type	Mic Trac	As Found	0.40007	Accuracy	0.00007	Fail Before	No
		As Left	0.40007	Accuracy	0.00007	Fail After	No

Gage ID of Standard	REF-G141308	Gage S/N	G141308	NIST No.	747V66-240502-00
Std Due Date	10/1/2026				

Standard ID	.50000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.49900	Nominal	0.50000	Maximum	0.50100
Ref Type	Mic Trac	As Found	0.50007	Accuracy	0.00007	Fail Before	No
		As Left	0.50007	Accuracy	0.00007	Fail After	No

Gage ID of Standard	REF-G141308	Gage S/N	G141308	NIST No.	747V66-240502-00
Std Due Date	10/1/2026				

Calibrated By Taichi Daimo Signature  Date: 5/5/2026

APPROVED
By Kayla Myers at 8:09 am, May 08, 2026