

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-05052026-017

<p>Gage ID PEA-J101486 Gage S/N J101486 Description Lead Gauge Operating Procedure: OP-124 Unit of Meas. inch Manufacturer Gagemaker Cal. Date 5/5/2026 Next Due 11/5/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	+0.00500"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.00450	Nominal	0.00500	Maximum	0.00550
Ref Type	Mic Trac	As Found	0.00525	Accuracy	0.00025	Fail Before	No
		As Left	0.00525	Accuracy	0.00025	Fail After	No
Gage ID of Standard	REF-G141308						
Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00		

Standard ID	+0.01000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.00950	Nominal	0.01000	Maximum	0.01050
Ref Type	Mic Trac	As Found	0.01041	Accuracy	0.00041	Fail Before	No
		As Left	0.01041	Accuracy	0.00041	Fail After	No
Gage ID of Standard	REF-G141308						
Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00		

Standard ID	-0.00500"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	-0.00550	Nominal	-0.00500	Maximum	-0.00450

Ref Type	Mic Trac	As Found	-0.00527	Accuracy	-0.00027	Fail Before	No
		As Left	-0.00527	Accuracy	-0.00027	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00
--------------	-----------	----------	---------	----------	------------------

Standard ID	-0.01000"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	-0.01050	Nominal	-0.01000	Maximum	-0.00950
Ref Type	Mic Trac	As Found	-0.01041	Accuracy	-0.00041	Fail Before	No
		As Left	-0.01041	Accuracy	-0.00041	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date	10/1/2026	Gage S/N	G141308	NIST No.	747V66-240502-00
--------------	-----------	----------	---------	----------	------------------

Calibrated By Taichi Daimo Signature  Date: 5/5/2026

APPROVED
By Kayla Myers at 8:07 am, May 08, 2026