

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-05202026-000

<p>Gage ID PEA-D13597 Gage S/N D13597 Description Lead Gauge Standard Operating Procedure: OP-125 Unit of Meas. Manufacturer Gagemaker Cal. Date 5/18/2026 Next Due 5/18/2029 Cal. Freq. 3.00 Years Location Lab</p>	<p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p align="center">Approved Yes Customer Info. Peak NDT</p>
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Gage: LS-1007 4 TPI, 2 TPF

Distance Groove #1-#2

Master	Actual	Deviation
+0.501730	+0.501760	+0.000030

Tol +0.000100 / -0.000100

Distance Remaining Grooves

Master	Actual	Deviation
+1.003470	+1.003500	+0.000040
+2.006930	+2.007000	+0.000070
+3.010400	+3.010450	+0.000060
+4.013860	+4.013930	+0.000070

Tol +0.000200 / -0.000200

Procedure # SDM-51 Rev.1
 Contact Point Size .144"

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

Reference Standard Serial #: K10567

Mic Trac (.48"), Gagemaker, Model MT4048
Traceable to NIST via NAMAS certification no: Cert# 176480
Last / Next Cal Dates: 12/05/2025 --> 12/05/2026

Calibrated By Taichi Daimo Signature  Date: 5/18/2026

APPROVED
By Kayla Myers at 12:22 pm, May 22, 2026