

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-05052026-019

<p><b>Gage ID</b> PEA-BR11GM0039  <b>Gage S/N</b> BR11GM0039  <b>Description</b> pitch Diameter Gauge  <b>Operating Procedure:</b> OP-122  <b>Unit of Meas.</b> inch  <b>Manufacturer</b> Gagemaker  <b>Cal. Date</b> 5/5/2026  <b>Next Due</b> 11/5/2026  <b>Cal. Freq.</b> 6.00                      Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>  <b>Temperature</b> 68 +/- 2 deg F  <b>Humidity</b> 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT</p>
--	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

---

<b>Standard ID</b> .10000 "	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b> 0.09900	<b>Nominal</b> 0.10000	<b>Maximum</b> 0.10100
<b>Ref Type</b> Mic Trac	<b>As Found</b> 0.09994	<b>Accuracy</b> -0.00006	<b>Fail Before</b> No
	<b>As Left</b> 0.09994	<b>Accuracy</b> -0.00006	<b>Fail After</b> No
<b>Gage ID of Standard</b> REF-G141308	<b>Gage S/N</b> G141308	<b>NIST No.</b> 747V66-240502-00	
<b>Std Due Date</b> 10/1/2026			

<b>Standard ID</b> .20000 "	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b> 0.19900	<b>Nominal</b> 0.20000	<b>Maximum</b> 0.20100
<b>Ref Type</b> Mic Trac	<b>As Found</b> 0.19994	<b>Accuracy</b> -0.00006	<b>Fail Before</b> No
	<b>As Left</b> 0.19994	<b>Accuracy</b> -0.00006	<b>Fail After</b> No
<b>Gage ID of Standard</b> REF-G141308	<b>Gage S/N</b> G141308	<b>NIST No.</b> 747V66-240502-00	
<b>Std Due Date</b> 10/1/2026			

<b>Standard ID</b> .30000 "	<b>Uncertainty</b>	<b>Units</b>	<b>Type</b> V
<b>Limited Use?</b> No	<b>Minimum</b> 0.29900	<b>Nominal</b> 0.30000	<b>Maximum</b> 0.30100

Ref Type	Mic Trac	As Found	0.30007	Accuracy	0.00007	Fail Before	No
		As Left	0.30007	Accuracy	0.00007	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 10/1/2026 Gage S/N G141308 NIST No. 747V66-240502-00

Standard ID	.40000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.39900	Nominal	0.40000	Maximum	0.40100
Ref Type	Mic Trac	As Found	0.40005	Accuracy	0.00005	Fail Before	No
		As Left	0.40005	Accuracy	0.00005	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 10/1/2026 Gage S/N G141308 NIST No. 747V66-240502-00

Standard ID	.50000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.49900	Nominal	0.50000	Maximum	0.50100
Ref Type	Mic Trac	As Found	0.50004	Accuracy	0.00004	Fail Before	No
		As Left	0.50004	Accuracy	0.00004	Fail After	No

Gage ID of Standard REF-G141308

Std Due Date 10/1/2026 Gage S/N G141308 NIST No. 747V66-240502-00

Calibrated By Taichi Daimo Signature  Date: 5/5/2026

**APPROVED**  
By Kayla Myers at 8:11 am, May 08, 2026