

1115 Garber Road  
Broussard LA 70518

**Company** Peak NDT Solutions  
**Address** 515A Verot School Road  
Lafayette LA  
**Zip** 70506  
**Page** 1 of 1

**Doc #:** 184652  
**OEM #:**  
**Contact**  
**Dept** WO# 061726-01  
**Cal Date** 06/17/2026  
**PO#:**

**Print#:**  
**Serial#:** 774-1  
**Location** Calibration Lab  
**Mfg.:** King  
**Model** Scope  
**Gage** 0-6mm

**Scope Readings As Found**

| Master    | Actual    | Deviation | Serial | Tol +0.100000 / -0.100000 |
|-----------|-----------|-----------|--------|---------------------------|
| +0.000000 | +0.000000 | +0.000000 |        |                           |
| +1.000000 | +1.000000 | +0.000000 |        |                           |
| +2.000000 | +2.000000 | +0.000000 |        |                           |
| +3.000000 | +3.000000 | +0.000000 |        |                           |
| +4.000000 | +4.000000 | +0.000000 |        |                           |
| +5.000000 | +5.000000 | +0.000000 |        |                           |
| +6.000000 | +6.000000 | +0.000000 |        |                           |

**Comments**

Procedure# SDM-12 Rev.1

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 17656

Stage Micrometer, Detroit Testing Model N/A

Traceable to NIST via NAMA certification no: Cert#

CQ047-42488-775

Last / Next Cal Dates: 02/16/2026 -> 02/16/2029

**Gage Status PASS**

**Next Calibration Due: 06/17/2027**

Certified By: Mason Pellerin

Signature 

This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revisions of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

**APPROVED**  
By Kayla Myers at 1:49 pm, Jun 22, 2026