

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-05052026-010

| | |
|--|--|
| <p>Gage ID PEA-16-2102 Gage S/N 16-2102 Description 4 Step Block Operating Procedure: OP-126 Unit of Meas. inch Manufacturer N/A Cal. Date 5/5/2026 Next Due 5/5/2027 Cal. Freq. 1.00 Years Location Lab</p> | <p>Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p> |
|--|--|

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

| | | | | | |
|---|--------------------------|-----------------|-----------------|-------|-----------------------|
| Standard ID .250 " | Uncertainty | | Units | | Type V |
| Limited Use? No | Minimum | 0.248 | Nominal | 0.250 | Maximum 0.252 |
| Ref Type 6" Digital Caliper | As Found | 0.250 | Accuracy | 0.000 | Fail Before No |
| | As Left | 0.250 | Accuracy | 0.000 | Fail After No |
| Gage ID of Standard REF-08341639 | Gage S/N 08341639 | NIST No. | CN304-48449-711 | | |
| Std Due Date 10/13/2026 | | | | | |

| | | | | | |
|---|--------------------------|-----------------|-----------------|-------|-----------------------|
| Standard ID .500 " | Uncertainty | | Units | | Type V |
| Limited Use? No | Minimum | 0.498 | Nominal | 0.500 | Maximum 0.502 |
| Ref Type 6" Digital Caliper | As Found | 0.500 | Accuracy | 0.000 | Fail Before No |
| | As Left | 0.500 | Accuracy | 0.000 | Fail After No |
| Gage ID of Standard REF-08341639 | Gage S/N 08341639 | NIST No. | CN304-48449-711 | | |
| Std Due Date 10/13/2026 | | | | | |

| | | | | | |
|---------------------------|--------------------|-------|----------------|-------|----------------------|
| Standard ID .750 " | Uncertainty | | Units | | Type V |
| Limited Use? No | Minimum | 0.748 | Nominal | 0.750 | Maximum 0.752 |

| | | | | | | | |
|----------|--------------------|----------|-------|----------|-------|-------------|----|
| Ref Type | 6" Digital Caliper | As Found | 0.750 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 0.750 | Accuracy | 0.000 | Fail After | No |

Gage ID of Standard REF-08341639

| | | | | | |
|--------------|------------|----------|----------|----------|-----------------|
| Std Due Date | 10/13/2026 | Gage S/N | 08341639 | NIST No. | CN304-48449-711 |
|--------------|------------|----------|----------|----------|-----------------|

| | | | | | | | |
|--------------|--------------------|-------------|-------|----------|-------|-------------|-------|
| Standard ID | 1.000 " | Uncertainty | | Units | | Type | V |
| Limited Use? | No | Minimum | 0.998 | Nominal | 1.000 | Maximum | 1.002 |
| Ref Type | 6" Digital Caliper | As Found | 1.000 | Accuracy | 0.000 | Fail Before | No |
| | | As Left | 1.000 | Accuracy | 0.000 | Fail After | No |

Gage ID of Standard REF-08341639

| | | | | | |
|--------------|------------|----------|----------|----------|-----------------|
| Std Due Date | 10/13/2026 | Gage S/N | 08341639 | NIST No. | CN304-48449-711 |
|--------------|------------|----------|----------|----------|-----------------|

Calibrated By Taichi Daimo Signature  Date: 5/5/2026

APPROVED
By Kayla Myers at 7:59 am, May 08, 2026