# TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

## Certificate No TSP-03012024-009

Gage ID PEA-28596

Gage S/N 28596

Description 10 lb. Test Weight

Operating Procedure:

OP-118

Unit of Meas. Ibs

Manufacturer Parker

Cal. Date 3/1/2024 Next Due 3/1/2027

Cal. Freq. 3.00

Years

Location Lab

**Environmental Conditions** 

Temperature

68 +/- 2 deg F

Humidity

20-55%

Approved Yes Customer Info. Peak NDT

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

## **Findings**

SO 12644 JOB 13263

AS FOUND: PASSED

AS LEFT: PASSED

Using calibrated scale serial no. 7129341906

Test weight as found 10.06 lbs.

Test weight as left 10.06 lbs.

Calibration due date for scale is on 4/13/2024.

Scale is traceable to NIST by the following numbers:LA22-255, LA22-256, LA22-257

Calibrated By Taichi Daimo

Signature

Date:

3/1/2024

# **APPROVED**

By Kayla Myers at 2:10 pm, Mar 04, 2024