

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-06162025-011

Gage ID PEA-CES-105679 Gage S/N CES-105679 Description 10 1/2 In. DC Magnetic Coil Operating Procedure: OP-116 Unit of Meas. Amps Manufacturer CES-Cut-N-Edge Service Cal. Date 6/16/2025 Next Due 12/16/2025 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

AS FOUND: PASSED AS LEFT: PASSED

Mag coil putting out magnetic field from both the pos. and neg. sides of coil equal to or greater than 50 gauss for calibration.

Calibrated 50-0-50 gauss meter serial no. 09-1234 was used for this calibration.

Calibration due date for gauss meter is on 05/12/26

Gauss meter is traceable to NIST by the following number: 1114538

Calibrated By Taichi Daimo Signature  Date: 6/16/2025

APPROVED*By Kayla Myers at 12:16 pm, Jun 17, 2025*