

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-12222025-018

Gage ID PEA-2324001545 Gage S/N 2324001545 Description UV Light Meter Operating Procedure: OP-119 Unit of Meas. Microwatts/cm2 Manufacturer General Cal. Date 12/22/2025 Next Due 6/22/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Calibrated black light serial no. 2234240 was also used during this calibration.
Calibration due date for black light is on 06/17/2026

Black light is traceable to NIST by the following numbers: 263546, 263547

Standard ID	12"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6678	Nominal	7029	Maximum	7380
Ref Type	Long Wave	As Found	7177	Accuracy	148	Fail Before	No
		As Left	7177	Accuracy	148	Fail After	No
Gage ID of Standard	REF-1944278						
Std Due Date	6/17/2026	Gage S/N	1944278/1944280	NIST No.	263546, 263547		

Standard ID	15"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	4564	Nominal	4804	Maximum	5044
Ref Type	Long Wave	As Found	4945	Accuracy	141	Fail Before	No
		As Left	4945	Accuracy	141	Fail After	No
Gage ID of Standard	REF-1944278						
Std Due Date	6/17/2026	Gage S/N	1944278/1944280	NIST No.	263546, 263547		

Calibrated By Taichi Daimo

Signature



Date: 12/22/2025

APPROVED

By Kayla Myers at 11:07 am, Dec 22, 2025