Gage ID: PEA-2324001545 Page 1 of 2

## **TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

### **Certificate of Calibration**

#### Certificate No TSP-06162025-010

Gage S/N 2324001545
Gage S/N 2324001545
Description UV Light Meter

Operating OP-119 Procedure:

Unit of Meas. Microwatts/cm2

Manufacturer General Cal. Date 6/16/2025 Next Due 12/16/2025

Cal. Freq. 6.00 Months

Location Lab

**Environmental Conditions** 

Temperature 68 +/- 2 deg F Humidity 20-55%

Approved Yes
Customer Info. Peak NDT

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### **Findings**

Calibrated black light serial no. 2234240 was also used during this calibration. Calibration due date for black light is on 12/13/2025

Black light is traceable to NIST by the following numbers: 258648, 258649

Standard ID	12"		Uncertainty		Units		Туре	V
Limited Use?	No		Minimum	6536	Nominal	6879	Maximum	7222
Ref Type	Long Wave		As Found	7033	Accuracy	154	Fail Before	No
			As Left	7033	Accuracy	154	Fail After	No
Gage ID of Sta	ndard	REF-1944278						
Std Due Date		6/17/2025	Gage S/N	1944278/1944 280	NIST No.	258648, 258649, 258650		
Standard ID	15"		Uncertainty	····	Units		Туре	٧
Limited Use?	No		Minimum	4419	Nominal	4651	Maximum	4883
Ref Type	Long	Wave	As Found	4798	Accuracy	147	Fail Before	No
			As Left	4798	Accuracy	147	Fail After	No
Gage ID of Sta	ndard	REF-1944278						
Std Due Date		6/17/2025	Gage S/N	1944278/1944 280	NIST No.	258648, 25864	9, 258650	

Gage ID: PEA-2324001545

Page 2 of 2

Calibrated By Taichi Daimo

Signature C

Date:

6/16/2025

# **APPROVED**

By Kayla Myers at 12:14 pm, Jun 17, 2025