1 of 1

TSP Calibration, Inc.

1110 S. FM 1788 #42 Phone: (432) 296-4992 Email: Chad@tspndt.com Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-11052025-018

Gage ID PEA-N7363

Gage S/N N7363

Description A/C Yoke

Operating

OP-115

Procedure:

Unit of Meas. pounds

Manufacturer Magnaflux

Cal. Date 11/5/2025 Next Due 5/5/2026

Cal. Freq. 6.00

Months

Location Lab

Environmental Conditions

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15104 Job 14863

As Found: Passed

As Left: Passed

A/C Yoke picked up and held required 10 pound weight for calibration.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):

Reference Standard: Test Weight - Serial Number: 3452

NIST Number: Doc#171848 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams

Signature

Date:

11/5/2025

APPROVED

By Kayla Myers at 3:56 pm, Nov 12, 2025