TSP Calibration, Inc.

1110 S. FM 1788 #42
Phone: (432) 296-4992 Email: Chad@tspndt.com
Midland,TX 79706

Certificate of Calibration

Certificate No TSPW-11052025-012

Gage ID PEA-9-622-1

Gage S/N 9-622-1

Description 20-0-20 gauss meter

Operating

OP-117

Procedure:

16.

Unit of Meas. gauss

Manufacturer R.B. Annis

Cal. Date 11/5/2025

Next Due 5/5/2026

Cal. Freq. 6.00

Months

Location Lab

Environmental Conditions

Temperature

68 Deg F +/- 2

Humidity

35-55%

Approved Yes

Customer Info. Peak NDT Solutions

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15104 Job 14863

imited Use?	No		Minimum		9	Nominal		10	Maximum	11	
Standard ID	-10		Uncertainty			Units	_	-	Туре	V	
Std Due Date		8/21/2026	Gage S/N	TSPSTD-02	NIST No.		171844				
Gage ID of Star	ndard	REF-TSPSTD-02									
			As Left		20	Accuracy		0	Fail After	No	
Ref Type 20-0		20 Gauss Meter	As Found		20	Accuracy		0	Fail Before	No	
Limited Use?	No		Minimum		19	Nominal		20	Maximum	21	
Standard ID	+20		Uncertainty			Units			Туре	· V	
old Due Dale		0/21/2020	Gage S/N	TSPSTD-02		NIST No.	171844				
Std Due Date	i dui u	8/21/2026	Carra 8/N	TODOTO 00							
Gage ID of Star	ndard	REF-TSPSTD-02				•					
			As Left		10	Accuracy		0	Fail After	No	
Ref Type	20-0-20 Gauss Meter		As Found		10	Accuracy		0	Fail Before	No.	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	
Standard ID	+10		Uncertainty			Units			Туре	V	

Page 2 of 2 Ref Type 20-0-20 Gauss Meter As Found 10 Accuracy Fail Before No As Left 10 Accuracy Fail After No Gage ID of Standard REF-TSPSTD-02 Std Due Date 8/21/2026 Gage S/N TSPSTD-02 NIST No. 171844 Standard ID -20 Uncertainty Units Type V Limited Use? No Minimum 19 Nominal 20 Maximum 21 Ref Type 20-0-20 Gauss Meter As Found 20 Accuracy 0 Fail Before No As Left 20 **Accuracy** 0 Fail After No Gage ID of Standard REF-TSPSTD-02 Std Due Date 8/21/2026 Gage S/N TSPSTD-02 NIST No. 171844

Date:

11/5/2025

APPROVED

Gage ID: PEA-9-622-1

Calibrated By Chad Adams

By Kayla Myers at 4:29 pm, Nov 11, 2025

Signature