

### TSP Calibration, Inc.

1110 S. FM 1788 #42  
Phone: (432) 296-4992 Email: Chad@tspndt.com  
Midland, TX 79706

## Certificate of Calibration

Certificate No TSPW-04172026-014

<p><b>Gage ID</b> PEA-CES-105588  <b>Gage S/N</b> CES-105588  <b>Description</b> 10-1/2" Magnetic Coil  <b>Operating Procedure:</b> OP-116  <b>Unit of Meas.</b> Gauss  <b>Manufacturer</b> Cut-N-Edge  <b>Cal. Date</b> 4/17/2026  <b>Next Due</b> 10/17/2026  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>  <b>Temperature</b> 68 Deg F +/- 2  <b>Humidity</b> 35-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT Solutions</p>
--	--

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

SO 15868 Job 15253

As Found: Passed As Left: Passed

Coil putting out magnetic field from both the positive and negative side equal to or greater than 50 gauss for calibration.

We certify the equipment used for this calibration is traceable to NIST through the following number(s):  
Reference Standard: 50-0-50 Magnetometer - Serial Number: TSPSTD-01  
NIST Number: Doc#171843 - Calibration Due Date: 8/21/2026

Calibrated By Chad Adams Signature  Date: 4/17/2026

**APPROVED**  
By Kayla Myers at 9:55 pm, Jun 02, 2026