

TSP Calibration, Inc.

1110 S. FM 1788 #42
 Phone: (432) 296-4992 Email: Chad@tspndt.com
 Midland, TX 79706

Certificate of Calibration

Certificate No TSPW-04172026-011

<p>Gage ID PEA-22261 Gage S/N 22261 Description UT Thickness Gauge Operating Procedure: OP-121 Unit of Meas. Inch Manufacturer Dakota Cal. Date 4/17/2026 Next Due 10/17/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions Temperature 68 Deg F +/- 2 Humidity 35-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT Solutions</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

SO 15868 Job 15253

Standard ID .100 "	Uncertainty		Units	Type V
Limited Use? No	Minimum	0.098	Nominal	Maximum 0.102
Ref Type 5 Step Block	As Found	0.100	Accuracy	Fail Before No
	As Left	0.100	Accuracy	Fail After No
Gage ID of Standard REF-337214	Gage S/N 337214	NIST No.	171846	
Std Due Date 8/21/2026				

Standard ID .200 "	Uncertainty		Units	Type V
Limited Use? No	Minimum	0.198	Nominal	Maximum 0.202
Ref Type 5 Step Block	As Found	0.200	Accuracy	Fail Before No
	As Left	0.200	Accuracy	Fail After No
Gage ID of Standard REF-337214	Gage S/N 337214	NIST No.	171846	
Std Due Date 8/21/2026				

Standard ID .300 "	Uncertainty		Units	Type V
Limited Use? No	Minimum	0.298	Nominal	Maximum 0.302

Ref Type	5 Step Block	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No


Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Standard ID	.400 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	5 Step Block	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Standard ID	.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	5 Step Block	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-337214	Gage S/N	337214	NIST No.	171846
Std Due Date	8/21/2026				

Calibrated By Chad Adams Signature  Date: 4/17/2026

APPROVED
By Kayla Myers at 9:53 pm, Jun 02, 2026