## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

### **Certificate of Calibration**

#### Certificate No TSP-02202025-034

Gage ID PEA-13839-6

Gage S/N 13839-6

**Description** 20-0-20 Gauss Meter

Operating OP-117 Procedure:

Unit of Meas. Gauss

Manufacturer R. B. Annis Co., Inc.

Cal. Date 2/20/2025 Next Due 8/20/2025

Cal. Freq. 6.00

Months

Location Lab

**Environmental Conditions** 

Temperature

68 +/- 2 deg F

Humidity

20-55%

**Approved** Yes Customer Info. Peak NDT

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### **Findings**

Standard ID	+10	Uncertainty			Units			Туре	V
Limited Use?	No	Minimum		9	Nominal		10	Maximum	11
Ref Type	Gauss Meter	As Found		10	Accuracy		0	Fail Before	No
		As Left		10	Accuracy		0	Fail After	No
Gage ID of Star	ndard Ref-09-1229								
Std Due Date	10/10/2025	Gage S/N	09-1229	NIST No.		114537			
Standard ID	+20	Uncertainty			Units			Туре	V
Limited Use?	No	Minimum		19	Nominal		20	Maximum	21
Ref Type	Gauss Meter	As Found		20	Accuracy		0	Fail Before	No
		As Left		20	Accuracy		0	Fail After	No
Gage ID of Star	ndard Ref-09-1229								
Std Due Date	10/10/2025	Gage S/N	09-1229	NIST No.		114537			
Standard ID	-10	Uncertainty			Units			Туре	V
Limited Use?	No	Minimum		9	Nominal		10	Maximum	11

Gage ID: PEA-13839-6

Page

2 of 2

Ref Type

Gauss Meter

As Found

10 Accuracy 0 Fail Before No

As Left

10 Accuracy 0 Fail After No

Gage ID of Standard

Ref-09-1229

Std Due Date

Ref Type

Std Due Date

10/10/2025

Gage S/N 09-1229

NIST No.

114537

Standard ID -20

Limited Use?

No

Gauss Meter

Uncertainty Minimum

As Left

As Found

19 20

Units Nominal Accuracy

Accuracy

Type V 20 Maximum 21 0 Fail Before No

0 Fail After No

Gage ID of Standard

Ref-09-1229

10/10/2025

Gage S/N 09-1229

NIST No.

20

114537

Calibrated By Taichi Daimo

Signature 5

Date:

2/20/2025

# **APPROVED**

By Kayla Myers at 1:49 pm, Feb 21, 2025