Gage ID: PEA-13897-6 Page 1 of 2

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-05212025-000

Gage ID PEA-13897-6 **Gage S/N** 13897-6

Description 20-0-20 Gauss Meter

Operating OP-117
Procedure:

Unit of Meas. Gauss

Manufacturer R. B. Annis Co., Inc.

Cal. Date 5/21/2025 Next Due 11/21/2025

Cal. Freq. 6.00 Months

Location Lab

Environmental Conditions

Temperature 68 +/- 2 deg F Humidity 20-55%

Approved Yes
Customer Info. Peak NDT

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	+10		Uncertainty			Units			Туре	٧	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	
Ref Type	Gauss	s Meter	As Found		10	Accuracy		0	Fail Before	No	
			As Left		10	Accuracy		0	Fail After	No	
Gage ID of Star	ndard	Ref-09-1229									
Std Due Date		10/10/2025	Gage S/N	09-1229	NIST No.		114537				
Standard ID	+20		Uncertainty			Units			Туре	v	
Limited Use?	No		Minimum		19	Nominal		20	Maximum	21	
Ref Type	Gauss	s Meter	As Found		20	Accuracy		0	Fail Before	No	
			As Left		20	Accuracy		0	Fail After	No	
Gage ID of Standard Ref-09-1229											
Std Due Date		10/10/2025	Gage S/N	09-1229	NIST No.		114537				
Standard ID	-10		Uncertainty			Units			Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	

Calibrated E	By T	aichi Daimo	Signature		1	10			Date:	5/21/2025
Gage ID of Star Std Due Date	dard Ref-09-1229 10/10/2025		Gage S/N	09-1229	NIST No.		114537			
			As Left		20	Accuracy		0	Fail After	No
Ref Type	Gauss	s Meter	As Found		20	Accuracy		0	Fail Before	No
Limited Use?	No		Minimum		19	Nominal		20	Maximum	21
Standard ID	-20		Uncertainty			Units			Туре	V
Std Due Date		10/10/2025	Gage S/N	09-1229	N	IIST No.	114537			
Gage ID of Star	ndard	Ref-09-1229								
			As Left		10	Accuracy		0	Fail After	No
Ref Type	Gauss	s Meter	As Found		10	Accuracy		0	Fail Before	No
	_					 Name of the state of the state				

Gage ID: PEA-13897-6

Page 2 of 2

APPROVED

By Kayla Myers at 1:57 pm, May 21, 2025