

**TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

**Certificate of Calibration**

Certificate No TSP-05212025-003

<b>Gage ID</b> PEA-220702209 <b>Gage S/N</b> 220702209 <b>Description</b> Visible Light Meter <b>Operating Procedure:</b> OP-130 <b>Unit of Meas.</b> Foot Candles <b>Manufacturer</b> General <b>Cal. Date</b> 5/21/2025 <b>Next Due</b> 11/21/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b>  Temperature 68 +/- 2 deg F Humidity 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
---	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

Using calibrated light meter serial no. 1944278 with sensor 1944280 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	96.2	97.7
Low	33.4	34.8

Readings As Left (foot candles)

	Master	Actual
High	96.2	97.7
Low	33.4	34.8

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater  
Cal due date for light meter is 06/17/2025  
NIST Traceable through : 258648, 258649, 258650

**APPROVED****By Kayla Myers at 2:03 pm, May 21, 2025**Calibrated By Taichi DaimoSignature Date: 5/21/2025