

Company	Peak NDT Solutions	Doc #:	186503	Print#:	
Address	515A Verot School Road	OEM #:		Serial#:	N6108
	Lafayette LA	Contact		Location	Calibration Lab
Zip:	70506	Dept	WO# 072726-14	Mfg.:	Magnaflux
Page	1 of 1	Cal Date	07/28/2026	Model	Yoke
		PO#:		Gage	115 VAC, 6.3 Amps, 50/60 Hz

Yoke Will Pick Up A 10 Pound Weight With Poles Spaced 2-4"

Master	Actual	Deviation	Serial	Tol
+10.000000	+10.000000	+0.000000		+0.000000 / 0.000000

Comments

Procedure# SDM-20 Rev.1

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 48XL

Weight Set 2oz-50 lbs, Rice Lake, Class "F" NIST

Traceable to NIST via NAMAS certification no: Cert# 25-334, 25-376, Last / Next Cal Dates 05/05/2025 -> 05/05/2028
25-335**Gage Status PASS****Next Calibration Due: 01/28/2027**

Certified By: Mason Pellerin

Signature



This certificate is not valid unless all 1 page(s) are present

NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions (68 Degrees F +/- 2 Degrees F, Humidity 20-55%)

APPROVED

By Kayla Myers at 12:31 pm, Aug 03, 2026