

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-01202026-000

<p>Gage ID PEA-N6108 Gage S/N N6108 Description A/C Yoke Operating Procedure: OP-115 Unit of Meas. lbs Manufacturer Magnaflux Cal. Date 1/20/2026 Next Due 7/20/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
--	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

AS FOUND: PASSED AS LEFT: PASSED

A/C Yoke picked up and held required 10 lb. test weight for calibration.
 Calibrated test weight serial no. 2253 was used for this calibration.
 Test weight calibration is due on 4/13/2026.
 Test weight is traceable to NIST by the following numbers: LA22-255, LA22-256, LA22-257

Calibrated By Taichi Daimo Signature  Date: 1/20/2026

APPROVED
 By Kayla Myers at 11:02 am, Jan 20, 2026