

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

### Certificate of Calibration

Certificate No TSP-04202026-023

<p><b>Gage ID</b> PEA-1076199  <b>Gage S/N</b> 1076199  <b>Description</b> 12" Digital Caliper  <b>Operating Procedure:</b> OP-110  <b>Unit of Meas.</b> Inch  <b>Manufacturer</b> Mitutoyo  <b>Cal. Date</b> 4/20/2026  <b>Next Due</b> 10/20/2026  <b>Cal. Freq.</b> 6.00                      Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b>                  Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT</p>
--	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

<b>Standard ID</b>	1.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type V</b>
<b>Limited Use?</b>	No	<b>Minimum</b>	0.999	<b>Nominal</b>	1.000	<b>Maximum</b> 1.001
<b>Ref Type</b>	GAGE BLOCKS	<b>As Found</b>	1.000	<b>Accuracy</b>	0.000	<b>Fail Before</b> No
		<b>As Left</b>	1.000	<b>Accuracy</b>	0.000	<b>Fail After</b> No
<b>Gage ID of Standard</b>	REF-GBS-061259	<b>Gage S/N</b>	061259	<b>NIST No.</b>	LMUA-131223 2018-3-20, AFS-276 2018-3-20	
<b>Std Due Date</b>	12/3/2027					

<b>Standard ID</b>	3.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type V</b>
<b>Limited Use?</b>	No	<b>Minimum</b>	2.999	<b>Nominal</b>	3.000	<b>Maximum</b> 3.001
<b>Ref Type</b>	GAGE BLOCKS	<b>As Found</b>	3.000	<b>Accuracy</b>	0.000	<b>Fail Before</b> No
		<b>As Left</b>	3.000	<b>Accuracy</b>	0.000	<b>Fail After</b> No
<b>Gage ID of Standard</b>	REF-GBS-061259	<b>Gage S/N</b>	061259	<b>NIST No.</b>	LMUA-131223 2018-3-20, AFS-276 2018-3-20	
<b>Std Due Date</b>	12/3/2027					

<b>Standard ID</b>	5.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type V</b>
<b>Limited Use?</b>	No	<b>Minimum</b>	4.999	<b>Nominal</b>	5.000	<b>Maximum</b> 5.001

Ref Type	GAGE BLOCKS	As Found	5.000	Accuracy	0.000	Fail Before	No
		As Left	5.000	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	GAGE BLOCKS	As Found	7.000	Accuracy	0.000	Fail Before	No
		As Left	7.000	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Standard ID	9.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	8.999	Nominal	9.000	Maximum	9.001
Ref Type	GAGE BLOCKS	As Found	9.000	Accuracy	0.000	Fail Before	No
		As Left	9.000	Accuracy	0.000	Fail After	No

Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223 2018-3-20, AFS-276 2018-3-20		

Calibrated By Taichi Daimo      Signature       Date: 4/20/2026

**APPROVED**  
By Kayla Myers at 11:59 am, Apr 21, 2026