

# TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

Certificate No TSP-04212025-034

<b>Gage ID</b> PEA-0026384 <b>Gage S/N</b> 0026384 <b>Description</b> 12" Digital Depth Gauge <b>Operating Procedure:</b> OP-123 <b>Unit of Meas.</b> inch <b>Manufacturer</b> Mitutoyo <b>Cal. Date</b> 4/21/2025 <b>Next Due</b> 10/21/2025 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b> <b>Temperature</b> 68 +/- 2 deg F <b>Humidity</b> 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
--	---

### Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

### Findings

<b>Standard ID</b>	1.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	0.999	<b>Nominal</b>	1.000	<b>Maximum</b>	1.001
<b>Ref Type</b>	GAGE BLOCKS	<b>As Found</b>	1.000	<b>Accuracy</b>	0.000	<b>Fail Before</b>	No
		<b>As Left</b>	1.000	<b>Accuracy</b>	0.000	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-GBS-061259						
<b>Std Due Date</b>	12/3/2027	<b>Gage S/N</b>	061259	<b>NIST No.</b>	LMUA-131223 2018-3-20, AFS-276 2018-3-20		
<b>Standard ID</b>	3.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	2.999	<b>Nominal</b>	3.000	<b>Maximum</b>	3.001
<b>Ref Type</b>	GAGE BLOCKS	<b>As Found</b>	3.000	<b>Accuracy</b>	0.000	<b>Fail Before</b>	No
		<b>As Left</b>	3.000	<b>Accuracy</b>	0.000	<b>Fail After</b>	No
<b>Gage ID of Standard</b>	REF-GBS-061259						
<b>Std Due Date</b>	12/3/2027	<b>Gage S/N</b>	061259	<b>NIST No.</b>	LMUA-131223 2018-3-20, AFS-276 2018-3-20		
<b>Standard ID</b>	5.000 "	<b>Uncertainty</b>		<b>Units</b>		<b>Type</b>	V
<b>Limited Use?</b>	No	<b>Minimum</b>	4.999	<b>Nominal</b>	5.000	<b>Maximum</b>	5.001

Ref Type	GAGE BLOCKS	As Found	4.999	Accuracy	-0.001	Fail Before	No
		As Left	4.999	Accuracy	-0.001	Fail After	No
Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223	2018-3-20, AFS-276	2018-3-20

Standard ID	7.000 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6.999	Nominal	7.000	Maximum	7.001
Ref Type	GAGE BLOCKS	As Found	6.999	Accuracy	-0.001	Fail Before	No
		As Left	6.999	Accuracy	-0.001	Fail After	No

Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223	2018-3-20, AFS-276	2018-3-20

Standard ID	9.000	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	8.999	Nominal	9.000	Maximum	9.001
Ref Type	GAGE BLOCKS	As Found	8.999	Accuracy	-0.001	Fail Before	No
		As Left	8.999	Accuracy	-0.001	Fail After	No

Gage ID of Standard	REF-GBS-061259						
Std Due Date	12/3/2027	Gage S/N	061259	NIST No.	LMUA-131223	2018-3-20, AFS-276	2018-3-20

Calibrated By	Taichi Daimo	Signature		Date:	4/21/2025
---------------	--------------	-----------	--	-------	-----------

**APPROVED***By Kayla Myers at 12:20 pm, Apr 22, 2025*