

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-10202025-021

Gage ID PEA-2092538 Gage S/N 2092538/2092541 Description Dual Light Meter Operating Procedure: OP-119/OP-130 Unit of Meas. Microwatts/cm2 & Foot Candles Manufacturer Spectroline Cal. Date 10/20/2025 Next Due 4/20/2026 Cal. Freq. 6.00 Months Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Light Meter s/n: 2092538
Sensor s/n: 2092541

Using calibrated light meter serial no. 1944278 with sensor 1944280 the following readings were taken.

Readings As Found (foot candles)

	Master	Actual
High	96.4	96.1
Low	35.3	34.8

Readings As Left (foot candles)

	Master	Actual
High	96.4	96.1
Low	35.3	34.8

Tolerance: +/- 5% or +/- 2 foot candles whichever is greater
Cal due date for light meter is 06/17/2026
NIST Traceable through : 263546, 263547

Calibrated black light serial no. 2234240 was also used during this calibration.
Calibration due date for black light is on 06/17/2026
Black light is traceable to NIST by the following numbers: 263546, 263547

Standard ID	12"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	6596	Nominal	6943	Maximum	7290
Ref Type	Long Wave	As Found	6761	Accuracy	-182	Fail Before	No
		As Left	6761	Accuracy	-182	Fail After	No

Gage ID of Standard REF-1944278

Std Due Date 6/17/2026

Gage S/N 1944278/1944 280 NIST No. 263546, 263547

Standard ID	15"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	4453	Nominal	4687	Maximum	4921
Ref Type	Long Wave	As Found	4505	Accuracy	-182	Fail Before	No
		As Left	4505	Accuracy	-182	Fail After	No

Gage ID of Standard REF-1944278

Std Due Date 6/17/2026

Gage S/N 1944278/1944 280 NIST No. 263546, 263547

Calibrated By Taichi Daimo

Signature



Date: 10/20/2025

APPROVED

By Kayla Myers at 9:07 am, Oct 21, 2025