Gage ID: PEA-8882 Page 1 of 2

## TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

### **Certificate of Calibration**

#### Certificate No TSP-10202025-017

Gage ID PEA-8882

Gage S/N 8882

**Description** 20-0-20 Gauss Meter

Operating OP-117 Procedure:

Unit of Meas. Gauss

Manufacturer R. B. Annis Co., Inc.

Cal. Date 10/20/2025 Next Due 4/20/2026

Cal. Freq. 6.00 Months

Location Lab

**Environmental Conditions** 

Temperature 68 +/- 2 deg F

Humidity 20-55%

> Approved Yes Customer Info. Peak NDT

#### **Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

#### **Findings**

Standard ID	+10		Uncertainty			Units			Туре	٧	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	
Ref Type	Gauss Meter		As Found		10	Accuracy		0	Fail Before	No	
			As Left		10	Accuracy		0	Fail After	No	
Gage ID of Star	ndard R	tef-09-1229									
Std Due Date	1	0/13/2026	Gage S/N	09-1229	NIST No.		114537				
Standard ID	+20		Uncertainty			Units			Туре	v	
Limited Use?	No		Minimum		19	Nominal		20	Maximum	21	
Ref Type	Gauss M	eter	As Found		19	Accuracy		-1	Fail Before	No	
			As Left		19	Accuracy		-1	Fail After	No	
Gage ID of Star	ndard R	tef-09-1229									
Std Due Date	10/13/2026		Gage S/N	09-1229	NIST No.		114537				
Standard ID	-10		Uncertainty			Units	<del></del>		Туре	V	
Limited Use?	No		Minimum		9	Nominal		10	Maximum	11	

Calibrated E	T	aichi Daimo	Signature		0	<u>'</u>			Date:	10/20/2025
Gage ID of Star Std Due Date	ndard	Ref-09-1229 10/13/2026	Gage S/N	09-1229	N	IST No.	114537			
			As Left		20	Accuracy		0	Fail After	No
Ref Type	Gauss	Meter	As Found		20	Accuracy		0	Fail Before	No
_imited Use?	No		Minimum		19	Nominal		20	Maximum	21
Standard ID	-20		Uncertainty			Units			Туре	V
Std Due Date		10/13/2026	Gage S/N	09-1229	N	IST No.	114537			
Gage ID of Standard		Ref-09-1229								
	101 0		As Left		10	Accuracy		0	Fail After	No
Ref Type	Gauss	Meter	As Found		10	Accuracy		0	Fail Before	
	_					A				

Page 2 of 2

# **APPROVED**

Gage ID: PEA-8882

By Kayla Myers at 9:04 am, Oct 21, 2025