

### TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518  
 Phone: (337) 236-6078 Email: taichi@tspndt.com

## Certificate of Calibration

Certificate No TSP-04202026-017

<p><b>Gage ID</b> PEA-CES-105519  <b>Gage S/N</b> CES-105519  <b>Description</b> 10 1/2 In. DC Magnetic Coil  <b>Operating Procedure:</b> OP-116  <b>Unit of Meas.</b> Amps  <b>Manufacturer</b> CES-Cut-N-Edge Service  <b>Cal. Date</b> 4/20/2026  <b>Next Due</b> 10/20/2026  <b>Cal. Freq.</b> 6.00 Months  <b>Location</b> Lab</p>	<p><b>Environmental Conditions</b></p> <p>Temperature 68 +/- 2 deg F                  Humidity 20-55%</p> <p style="text-align: center;"><b>Approved</b> Yes  <b>Customer Info.</b> Peak NDT</p>
---	--

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

AS FOUND: PASSED AS LEFT: PASSED

Mag coil putting out magnetic field from both the pos. and neg. sides of coil equal to or greater than 50 gauss for calibration.

Calibrated 50-0-50 gauss meter serial no. 09-1234 was used for this calibration.

Calibration due date for gauss meter is on 05/12/26

Gauss meter is traceable to NIST by the following number: 1114538

Calibrated By Taichi Daimo Signature  Date: 4/20/2026

**APPROVED**  
 By Kayla Myers at 11:54 am, Apr 21, 2026