

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-04142026-000

<p>Gage ID PEA-889 Gage S/N 889 Description Magnetic Particle Unit Operating Procedure: OP-120 Unit of Meas. Amps Manufacturer Parker Cal. Date 4/14/2026 Next Due 10/14/2026 Cal. Freq. 6.00 Months Location Lab</p>	<p>Environmental Conditions</p> <p>Temperature 68 +/- 2 deg F Humidity 20-55%</p> <p style="text-align: center;">Approved Yes Customer Info. Peak NDT</p>
---	--

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

AS FOUND: PASSED AS LEFT: PASSED

Gage: 115 Volts, 26 amps, 50/60 hertz, 750 max amps

Amp Readings As Found

Master	Actual	Deviation	Tol
+200	+200	+0	+/- 30
+310	+320	+10	+/- 46
+410	+420	+10	+/- 61
+600	+620	+20	+/- 90

Amp Readings As Left

Master	Actual	Deviation	Tol
+200	+200	+0	+/- 30
+310	+320	+10	+/- 46
+410	+420	+10	+/- 61
+600	+620	+20	+/- 90

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:

Reference Standard Serial #: 212031
 Shunt Meter Test Kit, Magnaflux, Model 621318
 NIST Traceable through: 936183, 936684, 920718

Last / Next Cal Dates: 08/04/2025 --> 08/31/2026

Calibrated By Taichi Daimo Signature  Date: 4/14/2026

APPROVED

By Kayla Myers at 10:13 am, Apr 14, 2026