

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518
Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-04212025-032

Gage ID PEA-H06431 Gage S/N H06431 Description 5 Step Block Operating Procedure: OP-126 Unit of Meas. inch Manufacturer N/A Cal. Date 4/21/2025 Next Due 4/21/2026 Cal. Freq. 1.00 Years Location Lab	Environmental Conditions Temperature 68 +/- 2 deg F Humidity 20-55% Approved Yes Customer Info. Peak NDT
--	---

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	0.100"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.098	Nominal	0.100	Maximum	0.102
Ref Type	6" Digital Caliper	As Found	0.100	Accuracy	0.000	Fail Before	No
		As Left	0.100	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-08341639						
Std Due Date	10/10/2025	Gage S/N	08341639	NIST No.	CN304-48449-711		
Standard ID	0.200"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.198	Nominal	0.200	Maximum	0.202
Ref Type	6" Digital Caliper	As Found	0.200	Accuracy	0.000	Fail Before	No
		As Left	0.200	Accuracy	0.000	Fail After	No
Gage ID of Standard	REF-08341639						
Std Due Date	10/10/2025	Gage S/N	08341639	NIST No.	CN304-48449-711		
Standard ID	0.300"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.298	Nominal	0.300	Maximum	0.302

Ref Type	6" Digital Caliper	As Found	0.300	Accuracy	0.000	Fail Before	No
		As Left	0.300	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/10/2025

Gage S/N 08341639

NIST No.

CN304-48449-711

Standard ID	0.400"	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.398	Nominal	0.400	Maximum	0.402
Ref Type	6" Digital Caliper	As Found	0.400	Accuracy	0.000	Fail Before	No
		As Left	0.400	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/10/2025

Gage S/N 08341639

NIST No.

CN304-48449-711

Standard ID	0.500 "	Uncertainty		Units		Type	V
Limited Use?	No	Minimum	0.498	Nominal	0.500	Maximum	0.502
Ref Type	6" Digital Caliper	As Found	0.500	Accuracy	0.000	Fail Before	No
		As Left	0.500	Accuracy	0.000	Fail After	No

Gage ID of Standard REF-08341639

Std Due Date 10/10/2025

Gage S/N 08341639

NIST No.

CN304-48449-711

Calibrated By Taichi Daimo

Signature



Date: 4/21/2025

APPROVED*By Kayla Myers at 12:19 pm, Apr 22, 2025*