

SDM CALIBRATION\$ LLC

1115 Garber Road
Broussard LA 70518

Calibration Report

Company	Peak NDT Solutions	Doc #:	187495	Print#:	
Address	515A Verot School Road Lafayette LA	OEM #:		Serial#:	0973-10
Zip:	70506	Contact		Location	Calibration Lab
Page	1 of 1	Dept	WO# 081726-01	Mfg.:	Drilco
		Cal. Date:	08/19/2026	Model	DC Coil
		PO#:		Gage	DC Coil

Comments

Procedure# SDM-17 Rev.1

NOTE: DC Coil went through a verification test using a +/- gauss magnetometer to insure that the unit functions properly.

We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers

Reference Standard Serial#: 15540250

Multimeter, Fluke Model 87V

Traceable to NIST via NAMAS certification no: Cert # 144311

Last / Next Cal Dates: 12/06/2023 -> 12/06/2026

Gage Status PASS

Next Calibration Due: 02/19/2027

Certified By: Mason Pellerin

Signature



This certificate is not valid unless all 1 page(s) are present

NOTE: IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS

- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement has not been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revision of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions: (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

APPROVED

By Kayla Myers at 3:26 pm, Aug 19, 2026