**SDM CALIBRATIONS LLC**

1115 Garber Road
Broussard LA 70518

# Calibration Report

**Company** Peak NDT Solutions
**Address** 515A Verot School Road
Lafayette LA
**Zip:** 70506
**Page:** 1 of 1

**Doc #:** 190858
**OEM #:**
**Contact**
**Dept:** WO# 102726-13
**Cal. Date:** 10/29/2026
**PO#:**

**Print#:**
**Serial#:** 1708221680
**Location** Calibration Lab
**Mfg.:** Insize
**Model** Digital Caliper
**Gage** 12"

## OD Jaw Values

| Master | Actual | Deviation | Serial | Tol +0.001500 / -0.001500 |
|---|---|---|---|---|
| +0.000000 | +0.000000 | +0.000000 | | |
| +0.500000 | +0.500000 | +0.000000 | | |
| +1.000000 | +0.999500 | -0.000500 | | |
| +6.000000 | +5.999500 | -0.000500 | | |
| +12.000000 | +11.999500 | -0.000500 | | |

## ID Jaw Value

| Master | Actual | Deviation | Serial | Tol +0.001500 / -0.001500 |
|---|---|---|---|---|
| +0.200000 | +0.199500 | -0.000500 | | |

## Depth Rod Value

| Master | Actual | Deviation | Serial | Tol +0.001500 / -0.001500 |
|---|---|---|---|---|
| +0.000000 | +0.000000 | +0.000000 | | |

## Comments

Procedure# SDM-44 Rev.1

*We certify the equipment used for this calibration is traceable to NIST through one or more of the following numbers:*

Reference Standard Serial#: 5.3155HD16
Gage Block Set 46pc., Starrett, Model Grade B
Traceable to NIST via NAMAS certification no: Cert# CN304-48449-711

Last / Next Cal Dates: 10/31/2023 --> 10/31/2026

**Gage Status PASS**

**Next Calibration Due: 04/29/2027**

Certified By: Mason Pellerin

Signature Mas Pell

This certificate is not valid unless all 1 page(s) are present
NOTE IF ONLY AN "AS FOUND" READING IS STATED "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS
- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement hasn't been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revisions of applicable standards/specifications unless otherwise specified by the customer.
Environmental Conditions (68 Degrees F +/- 2 Degrees) (Humidity 20-55%)

**APPROVED**
**By Kayla Myers at 1:39 pm, Oct 29, 2026**