Gage ID: PEA-41321 Page 1 of 2

TSP Calibration, Inc.

3501 US Hwy 90 E Broussard, LA 70518 Phone: (337) 236-6078 Email: taichi@tspndt.com

Certificate of Calibration

Certificate No TSP-08062025-003

Gage ID PEA-41321
Gage S/N 41321

Description 50-0-50 Gauss Meter

Operating OP-117

Procedure:

Unit of Meas. Gauss

Manufacturer R. B. Annis Co., Inc.

Cal. Date 8/6/2025 Next Due 2/6/2026

Cal. Freq. 6.00 Months

Location Lab

Environmental Conditions

Temperature 68 +/- 2 deg F

Humidity 20-55%

Approved Yes **Customer Info.** Peak NDT

Certification Statement

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

Findings

Standard ID	+25		Uncertainty			Units		Туре	V	
Limited Use?	No		Minimum		24	Nominal	2	25 Maximum	26	
Ref Type	Gauss Meter		As Found		25	Accuracy		0 Fail Before	No	
			As Left		25	Accuracy		0 Fail After	No	
Gage ID of Standard REF-09-1234										
Std Due Date		5/12/2026	Gage S/N 09-1234		NIST No.		1114538			
Standard ID	+50		Uncertainty			Units		Туре	V	
Limited Use?	No		Minimum		49	Nominal	5	0 Maximum	51	
Ref Type	Gauss	Meter	As Found		50	Accuracy		0 Fail Before	No	
			As Left		50	Accuracy		0 Fail After	No	
Gage ID of Standard REF-09-1234										
Std Due Date		5/12/2026	Gage S/N	09-1234	NIST No.		1114538			
Standard ID	-25	 	Uncertainty			Units		Туре	٧	
Limited Use?	No		Minimum		24	Nominal	2	5 Maximum	26	

Ref Type Gauss Meter As Found 26 Accuracy Fail Before No As Left 26 Accuracy Fail After No Gage ID of Standard REF-09-1234 Std Due Date 5/12/2026 Gage S/N 09-1234 NIST No. 1114538 Standard ID -50 Units Uncertainty Type V Limited Use? No Minimum 49 Nominal 50 Maximum 51 Ref Type Gauss Meter As Found 51 Accuracy 1 Fail Before No As Left 51 Accuracy 1 Fail After No Gage ID of Standard REF-09-1234 Std Due Date 5/12/2026 Gage S/N 09-1234 NIST No. 1114538 Signature

Page

Date:

2 of 2

8/6/2025

APPROVED

Gage ID: PEA-41321

Calibrated By Taichi Daimo

By Kayla Myers at 2:36 pm, Aug 06, 2025