

**TSP Calibration, Inc.**

3501 US Hwy 90 E Broussard, LA 70518  
Phone: (337) 236-6078 Email: taichi@tspndt.com

**Certificate of Calibration**

Certificate No TSP-08042025-008

<b>Gage ID</b> PEA-RUN045 <b>Gage S/N</b> RUN045 <b>Description</b> Lead Gauge <b>Operating Procedure:</b> OP-124 <b>Unit of Meas.</b> Inch <b>Manufacturer</b> Gagemaker <b>Cal. Date</b> 8/4/2025 <b>Next Due</b> 2/4/2026 <b>Cal. Freq.</b> 6.00 Months <b>Location</b> Lab	<b>Environmental Conditions</b>  Temperature 68 +/- 2 deg F Humidity 20-55%  <b>Approved</b> Yes <b>Customer Info.</b> Peak NDT
---	---

**Certification Statement**

TSP Calibration, Inc. calibration systems complies with the requirements of ISO 9001:2015. The equipment that is certified by this certificate has been calibrated by standards that have accuracy which is traceable to standards of the National Institute of Standards and Technology.

**Findings**

AS FOUND: PASSED AS LEFT: PASSED

Gage: 0-.010"

Measured Values As Found

Master	Actual	Deviation
+0.000000	+0.000000	+0.000000
+0.005000	+0.005030	+0.000030
+0.009000	+0.009060	+0.000060
-0.005000	-0.005040	-0.000040
-0.009000	-0.009060	-0.000060

Tol +0.000500 / -0.000500**APPROVED****By Kayla Myers at 9:57 am, Aug 05, 2025**

Procedure# SDM-39 Rev.1  
Reference Standard Serial #: K10567  
Mic Trac (.48"), Gagemaker, Model MT4048  
Traceable to NIST via NAMAS certification no: Cert# 160673  
Last / Next Cal Dates: 12/07/2024 --> 12/07/2025

Calibrated By Taichi Daimo

Signature



Date:

8/4/2025