SDM Calibrations
1115 Garber Rd
Broussard, LA 70518
337 - 837 - 9040

CALIBRATION REPORT

Company Name: Peak NDT Solutions
Location: 515A Verot School Rd., Lafayette LA 70508
WO#: 100726-02
PO#:
Calibrated: 10/07/2026

Calibration ID: 95733
Serial Number: 93342
Manufacturer : New Tech
Model #: Dial Indicator
Gage: 1" .001"Graduation

Calibration Specifications

Total Range Up/Down Scale

Nominal	Unit	As Found&As Left	Tolerance	Deviation	Pass/Fail
0.000000	Inch	0.000000	+0.001000/-0.001000	0.000000	Pass
0.100000	Inch	0.100369	+0.001000/-0.001000	0.000369	Pass
0.250000	Inch	0.249997	+0.002000/-0.002000	-0.000003	Pass
0.375000	Inch	0.375048	+0.002000/-0.002000	0.000048	Pass
0.500000	Inch	0.500690	+0.002000/-0.002000	0.000690	Pass
0.375000	Inch	0.375225	+0.002000/-0.002000	0.000225	Pass
0.250000	Inch	0.249879	+0.002000/-0.002000	-0.000121	Pass
0.100000	Inch	0.100322	+0.001000/-0.001000	0.000322	Pass
0.000000	Inch	0.000000	+0.001000/-0.001000	0.000000	Pass

Test Instruments Used During the Calibration

Description	Test Instrument ID	Last Cal Date	Next Cal Date	NIST Number
LabMaster (80")	LMUA-131223	12/17/2024	12/17/2027	Cert#LMUA-131223 2018-3-20,AFS-276 2018-3-20
Gage Block Set Grade 00	XLRA07	10/13/2025	10/13/2028	Cert# CP286-38963-711

Notes: Procedure # SDM-41 Rev.1

Next Cal Date: 10/07/2027

Calibration Result: **Pass**

Calibrated By: Nicholas Hebert

This certificate is not valid unless all 1 page(s) are present:

NOTE: IF ONLY AN "AS FOUND" READING IS STATED, "AS FOUND AND AS LEFT" READINGS ARE THE SAME WITH NO ADJUSTMENTS.
- SDM Calibrations certifies the instrument above conforms to the original manufacturer's specifications and has been calibrated using standards whose accuracies are traceable to the (SI) via the National Institute of Standards and Technology (NIST) or have been derived from accepted values of natural physical constants or by the ratio type of self calibration techniques. The uncertainty of measurement hasn't been applied as part of the decision rules for Pass/Fail status. Calibrations are performed to the latest revisions of applicable standards/specifications unless otherwise specified by the customer.

Environmental Conditions: (68 Degrees F +/- 2 Degrees)(Humidity 20-55%)

APPROVED
By Kayla Myers at 9:57 am, Oct 09, 2026